

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080818\
 Data File : VN050467.D
 Acq On : 8 Aug 2018 18:32
 Operator : MD\SY
 Sample : J4380-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELD-BLANK

Quant Time: Aug 09 09:38:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	804217	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1243550	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1075698	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	447885	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	513312	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
35) Dibromofluoromethane	7.59	113	469354	47.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.08%	
50) Toluene-d8	10.09	98	1694823	45.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.36%	
62) 4-Bromofluorobenzene	12.40	95	514561	40.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.48%	
Target Compounds						
16) Acetone	3.82	43	28211	8.46	ug/l	96
20) Methylene Chloride	4.55	84	302237	30.23	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080818\
Data File : VN050467.D
Acq On : 8 Aug 2018 18:32
Operator : MD\SY
Sample : J4380-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELD-BLANK

Quant Time: Aug 09 09:38:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

