

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080818\
 Data File : VN050468.D
 Acq On : 8 Aug 2018 18:57
 Operator : MD\SY
 Sample : J4380-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-37

Manual Integrations
 APPROVED

MMDadoda
 8/9/2018 1:32:16 PM

Quant Time: Aug 09 09:44:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	806596	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1248072	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1053237	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	438254	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	507941	47.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.26%	
35) Dibromofluoromethane	7.59	113	468062	46.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.48%	
50) Toluene-d8	10.09	98	1700421	45.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.32%	
62) 4-Bromofluorobenzene	12.40	95	498866	38.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.74%	
Target Compounds						
16) Acetone	3.82	43	22151	6.54	ug/l	93
19) Methyl tert-butyl Ether	5.05	73	59148	2.39	ug/l	99
22) Diisopropyl ether	5.95	45	36473m	1.27	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080818\
Data File : VN050468.D
Acq On : 8 Aug 2018 18:57
Operator : MD\SY
Sample : J4380-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-37

Manual Integrations
APPROVED

MMDadoda
8/9/2018 1:32:16 PM

Quant Time: Aug 09 09:44:57 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

