


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN080818\
 Data File : VN050450.D
 Acq On : 8 Aug 2018 10:02
 Operator : MD\SY
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Title : SW846 8260
 Last Update : Tue Aug 07 05:22:15 2018



AutoFind: Scans 3362, 3363, 3364; Background Corrected with Scan 3349

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.1	21784	PASS
75	95	30	60	48.9	62368	PASS
95	95	100	100	100.0	127613	PASS
96	95	5	9	6.9	8784	PASS
173	174	0.00	2	0.8	917	PASS
174	95	50	100	92.5	118018	PASS
175	174	5	9	7.8	9162	PASS
176	174	95	101	97.2	114762	PASS
177	176	5	9	6.5	7426	PASS