

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080819\
 Data File : VN057126.D
 Acq On : 8 Aug 2019 12:45
 Operator : JC/SP
 Sample : K4092-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-01-080719-C

Quant Time: Aug 09 03:40:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1535252	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	2151327	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.90	117	1812014	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.87	152	478129	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	713929	45.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.44%	
35) Dibromofluoromethane	6.99	113	643647	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
50) Toluene-d8	9.56	98	2516183	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
62) 4-Bromofluorobenzene	11.92	95	676047	40.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.24%	
Target Compounds						
16) Acetone	3.48	43	59901	12.562	ug/l	99
20) Methylene Chloride	4.10	84	107179	6.811	ug/l	94
43) Isopropyl Acetate	7.56	43	54875	2.202	ug/l #	90
95) Naphthalene	14.66	128	7674	3.425	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080819\
Data File : VN057126.D
Acq On : 8 Aug 2019 12:45
Operator : JC/SP
Sample : K4092-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EO-01-080719-C

Quant Time: Aug 09 03:40:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

