

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080819\
 Data File : VN057127.D
 Acq On : 8 Aug 2019 13:09
 Operator : JC/SP
 Sample : K4092-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-01-080719-B

Quant Time: Aug 09 03:42:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1488168	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	2115230	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.90	117	1772569	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.87	152	470818	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	697237	45.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.12%	
35) Dibromofluoromethane	6.99	113	634702	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
50) Toluene-d8	9.56	98	2451817	48.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.56%	
62) 4-Bromofluorobenzene	11.92	95	669012	40.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.78%	
Target Compounds						
16) Acetone	3.48	43	66866	14.467	ug/l	92
20) Methylene Chloride	4.10	84	1506699	98.773	ug/l	98
30) Chloroform	6.75	83	32385	1.268	ug/l	95
43) Isopropyl Acetate	7.57	43	70438	2.875	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080819\
Data File : VN057127.D
Acq On : 8 Aug 2019 13:09
Operator : JC/SP
Sample : K4092-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EO-01-080719-B

Quant Time: Aug 09 03:42:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

