

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080819\
 Data File : VN057130.D
 Acq On : 8 Aug 2019 14:22
 Operator : JC/SP
 Sample : K3613-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AU-05-080719-B

Manual Integrations
 APPROVED

MMDadoda
 8/9/2019 9:28:04 AM

Quant Time: Aug 09 03:55:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1463542	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	2061195	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.90	117	1724464	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.87	152	485964	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	672604	44.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.38%	
35) Dibromofluoromethane	6.99	113	622127	48.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.20%	
50) Toluene-d8	9.56	98	2420775	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	11.92	95	660215	41.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.82%	
Target Compounds						
16) Acetone	3.48	43	54204	11.925	ug/l	98
18) Methyl Acetate	3.90	43	30448	1.612	ug/l #	72
20) Methylene Chloride	4.10	84	198210	13.212	ug/l	96
43) Isopropyl Acetate	7.57	43	106770m	4.472	ug/l	
95) Naphthalene	14.67	128	24941	4.783	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080819\
Data File : VN057130.D
Acq On : 8 Aug 2019 14:22
Operator : JC/SP
Sample : K3613-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AU-05-080719-B

Manual Integrations
APPROVED

MMDadoda
8/9/2019 9:28:04 AM

Quant Time: Aug 09 03:55:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

