

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080822\
 Data File : VN073774.D
 Acq On : 08 Aug 2022 16:50
 Operator : JC\MD
 Sample : N4079-02 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EFF-WW

Quant Time: Aug 09 00:28:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072722W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 27 18:07:26 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/09/2022
 Supervised By : Mahesh Dadoda 08/09/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.586	128	49346	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.904	114	250996	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.686	117	231387	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.369	65	135483	31.755	ug/l	0.00
Spiked Amount 30.000	Range	91 - 110	Recovery	=	105.833%	
60) 4-Bromofluorobenzene	12.674	95	109685	26.772	ug/l	0.00
Spiked Amount 30.000	Range	63 - 112	Recovery	=	89.233%	
63) Toluene-d8	10.380	98	374828	28.769	ug/l	0.00
Spiked Amount 30.000	Range	91 - 112	Recovery	=	95.900%	
Target Compounds						
15) Acetone	4.222	58	1704048	2529.526	ug/l	Qvalue 91
18) Methylene Chloride	5.022	84	3727m	0.952	ug/l	
25) Chloroform	7.745	83	7033	1.030	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080822\
Data File : VN073774.D
Acq On : 08 Aug 2022 16:50
Operator : JC\MD
Sample : N4079-02 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EFF-WW

Quant Time: Aug 09 00:28:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072722W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jul 27 18:07:26 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/09/2022
Supervised By :Mahesh Dadoda 08/09/2022

