

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080822\
 Data File : VN073775.D
 Acq On : 08 Aug 2022 17:15
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Aug 09 00:29:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072722W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jul 27 18:07:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.587	128	50607	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.904	114	253609	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.686	117	240572	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.369	65	133996	30.624	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.067%
60) 4-Bromofluorobenzene	12.674	95	118068	27.718	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.400%
63) Toluene-d8	10.380	98	379121	27.987	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	93.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080822\
Data File : VN073775.D
Acq On : 08 Aug 2022 17:15
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Aug 09 00:29:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N072722W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jul 27 18:07:26 2022
Response via : Initial Calibration

