

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080823\
 Data File : VN078806.D
 Acq On : 08 Aug 2023 18:40
 Operator : JC\MD
 Sample : VSTDICCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC020

Quant Time: Aug 09 04:41:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 09 04:37:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	32248	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	185975	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	156856	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	79883	32.252	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 107.500%	
60) 4-Bromofluorobenzene	12.853	95	69391	29.197	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 97.333%	
63) Toluene-d8	10.571	98	237195	30.203	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 100.667%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.136	85	47191	22.009	ug/l	98
3) Chloromethane	2.371	50	56042	21.899	ug/l	98
4) Vinyl Chloride	2.524	62	51065	15.601	ug/l #	86
5) Bromomethane	2.971	94	37743	15.369	ug/l	87
6) Chloroethane	3.130	64	36553	16.703	ug/l	97
7) Trichlorofluoromethane	3.507	101	74203	20.729	ug/l	91
8) Diethyl Ether	3.977	74	25763	21.662	ug/l	77
9) 1,1,2-Trichlorotrifluo...	4.377	101	39793	20.541	ug/l	91
10) 1,1-Dichloroethene	4.348	96	37546	20.464	ug/l	90
11) Methyl Iodide	4.606	142	50473	19.608	ug/l	93
12) Methyl Acetate	5.042	43	82925	25.806	ug/l	95
13) Acrolein	4.195	56	40428	120.806	ug/l	96
14) Acrylonitrile	5.730	53	110203	119.944	ug/l	96
15) Acetone	4.442	58	28047	108.176	ug/l	77
16) Carbon Disulfide	4.724	76	111045	21.237	ug/l #	94
17) Allyl chloride	5.036	41	60646	23.581	ug/l	91
18) Methylene Chloride	5.283	84	47269	21.186	ug/l #	91
19) trans-1,2-Dichloroethene	5.795	96	42756	20.086	ug/l	93
20) Diisopropyl ether	6.683	45	146795	24.629	ug/l #	93
21) 1,1-Dichloroethane	6.577	63	83556	21.850	ug/l	99
22) cis-1,2-Dichloroethene	7.489	96	26194	12.188	ug/l #	23
23) tert-Butyl Alcohol	5.536	59	39298	108.207	ug/l #	100
24) Methyl tert-Butyl Ether	5.812	73	126258	20.428	ug/l	92
25) Chloroform	7.977	83	82652	20.761	ug/l	89
26) Cyclohexane	8.259	56	63808	22.807	ug/l #	98
29) 1,1-Dichloropropene	8.383	75	61681	21.685	ug/l	97
30) 2-Butanone	7.494	43	147386	125.723	ug/l	92
31) 2,2-Dichloropropane	7.494	77	63260	19.927	ug/l	97
32) 1,1,1-Trichloroethane	8.183	97	69151	19.784	ug/l	94
33) Carbon Tetrachloride	8.365	117	62426	19.973	ug/l	97
34) Benzene	8.612	78	182106	21.407	ug/l	95
35) Methacrylonitrile	7.789	41	34461	26.279	ug/l	85
36) 1,2-Dichloroethane	8.677	62	63887	21.497	ug/l	99
37) Trichloroethene	9.359	130	44449	19.029	ug/l	94
38) Methylcyclohexane	9.606	83	59023	20.334	ug/l	95
39) 1,2-Dichloropropane	9.624	63	49101	21.313	ug/l	89
40) Dibromomethane	9.712	93	31583	20.245	ug/l	98
41) Bromodichloromethane	9.894	83	66664	21.236	ug/l	98
42) Vinyl Acetate	6.618	43	469611	124.787	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080823\
 Data File : VN078806.D
 Acq On : 08 Aug 2023 18:40
 Operator : JC\MD
 Sample : VSTDICCC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1


Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC020


Quant Time: Aug 09 04:41:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 09 04:37:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.565	43	20493	8.304	ug/l #	31
44) Isopropyl Acetate	8.694	43	107010	25.258	ug/l #	66
45) 1,4-Dioxane	9.700	88	15982	422.840	ug/l #	80
46) Methyl methacrylate	9.688	41	46397	23.964	ug/l	91
47) n-amyl Acetate	12.500	43	79278	24.305	ug/l	95
48) t-1,3-Dichloropropene	10.841	75	66659	20.289	ug/l	96
49) cis-1,3-Dichloropropene	10.318	75	74589	20.938	ug/l #	90
50) 1,1,2-Trichloroethane	11.024	97	44160	20.349	ug/l	96
51) Ethyl methacrylate	10.883	69	60595	20.289	ug/l #	65
52) 1,3-Dichloropropane	11.171	76	75854	19.226	ug/l	99
53) Dibromochloromethane	11.365	129	44805	18.612	ug/l	97
54) 1,2-Dibromoethane	11.477	107	42686	19.726	ug/l	98
55) 2-Chloroethyl vinyl ether	10.165	63	112133	114.352	ug/l	97
56) Bromoform	12.582	173	28780	17.988	ug/l	97
58) 4-Methyl-2-Pentanone	10.453	43	315458	123.975	ug/l #	91
59) 2-Hexanone	11.200	43	221370	117.699	ug/l	91
61) Tetrachloroethene	11.112	164	36710	17.910	ug/l	96
62) Toluene	10.635	91	192478	20.064	ug/l	100
64) Chlorobenzene	11.900	112	111495	20.010	ug/l	96
65) 1,1,1,2-Tetrachloroethane	11.965	131	43056	19.736	ug/l	99
66) Ethyl Benzene	11.971	91	204995	21.081	ug/l	92
67) m/p-Xylenes	12.077	106	146780	39.958	ug/l	99
68) o-Xylene	12.406	106	74229	19.732	ug/l	98
69) Styrene	12.418	104	111990	18.611	ug/l	97
70) Isopropylbenzene	12.700	105	180232	19.479	ug/l	98
71) 1,1,2,2-Tetrachloroethane	12.941	83	61793	20.305	ug/l	99
72) 1,2,3-Trichloropropane	13.000	75	74030	29.879	ug/l	53
73) Bromobenzene	12.982	156	42693	18.949	ug/l	97
74) n-propylbenzene	13.041	91	204759	19.793	ug/l	99
75) 2-Chlorotoluene	13.129	91	127452	17.924	ug/l	97
76) 1,3,5-Trimethylbenzene	13.176	105	148955	18.013	ug/l	97
77) t-1,4-Dichloro-2-butene	12.741	75	17182	19.427	ug/l	78
78) 4-Chlorotoluene	13.229	91	119075	17.569	ug/l	97
79) tert-butylbenzene	13.441	119	123189	17.385	ug/l	98
80) 1,2,4-Trimethylbenzene	13.488	105	140130	17.861	ug/l	99
81) sec-Butylbenzene	13.624	105	164005	17.886	ug/l	100
82) p-Isopropyltoluene	13.735	119	127838	17.289	ug/l	99
83) 1,3-Dichlorobenzene	13.735	146	69911	17.601	ug/l	99
84) 1,4-Dichlorobenzene	13.818	146	65844	17.671	ug/l	99
85) n-Butylbenzene	14.059	91	92573	17.978	ug/l	98
86) Hexachloroethane	14.341	117	26063	19.046	ug/l	90
87) 1,2-Dichlorobenzene	14.112	146	68778	18.383	ug/l	96
88) 1,2-Dibromo-3-Chloropr...	14.723	75	8617	17.075	ug/l	98
89) 1,2,4-Trichlorobenzene	15.400	180	19027	17.981	ug/l	97
90) Hexachlorobutadiene	15.506	225	14045	17.647	ug/l	95
91) Naphthalene	15.647	128	58324	18.200	ug/l	97
92) 1,2,3-Trichlorobenzene	15.847	180	22358	18.530	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 09 04:41:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 09 04:37:34 2023
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 7.824 min Scan# 998
Delta R.T. -0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Tgt Ion:128 Resp: 32248
Ion Ratio Lower Upper
128 100
49 201.5 0.0 370.8
130 128.7 0.0 307.0






#2
Dichlorodifluoromethane
Concen: 22.009 ug/l
RT: 2.136 min Scan# 31
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 85 Resp: 47191
Ion Ratio Lower Upper
85 100
87 31.7 16.4 49.4



Abundance Scan 71 (2.371 min): VN078806.D\data.ms (-66)



#3

Chloromethane

Concen: 21.899 ug/l

RT: 2.371 min Scan# 71

Delta R.T. 0.000 min

Lab File: VN078806.D

Acq: 08 Aug 2023 18:40

Instrument :

MSVOA_N

ClientSampleId :


VSTDICCC020

Tgt Ion: 50 Resp: 56042




Ion Ratio Lower Upper

50 100

52 29.5 24.6 36.8



Abundance Scan 97 (2.524 min): VN078806.D\data.ms (-90)



#4

Vinyl Chloride

Concen: 15.601 ug/l

RT: 2.524 min Scan# 97

Delta R.T. 0.000 min

Lab File: VN078806.D


Acq: 08 Aug 2023 18:40


Tgt Ion: 62 Resp: 51065

Ion Ratio Lower Upper

62 100

64 36.5 23.3 34.9#









#5
Bromomethane
Concen: 15.369 ug/l
RT: 2.971 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 94 Resp: 37743
Ion Ratio Lower Upper
94 100
96 80.1 73.9 110.9



#6
Chloroethane
Concen: 16.703 ug/l
RT: 3.130 min Scan# 200
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 64 Resp: 36553
Ion Ratio Lower Upper
64 100
66 30.6 25.7 38.5









#7
Trichlorofluoromethane
Concen: 20.729 ug/l
RT: 3.507 min Scan# 264
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Tgt Ion:101 Resp: 74203
Ion Ratio Lower Upper
101 100
103 55.7 50.4 75.6




#8
Diethyl Ether
Concen: 21.662 ug/l
RT: 3.977 min Scan# 344
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 74 Resp: 25763
Ion Ratio Lower Upper
74 100
45 101.8 16.3 146.5




Abundance Scan 412 (4.377 min): VN078806.D\data.ms (-32)




m/z-->

Abundance Scan 412 (4.377 min): VN078806.D\data.ms




m/z-->

Abundance Scan 412 (4.377 min): VN078806.D\data.ms (-32)




m/z-->

Abundance Scan 407 (4.348 min): VN078806.D\data.ms (-32)




m/z-->

Abundance Scan 407 (4.348 min): VN078806.D\data.ms



m/z-->

Abundance Scan 407 (4.348 min): VN078806.D\data.ms (-32)




m/z-->

#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.541 ug/l
RT: 4.377 min Scan# 412
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40


Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:	101	Resp:	39793
Ion	Ratio	Lower	Upper
101	100		
85	35.5	0.0	89.2
151	69.6	0.0	148.8



#10
1,1-Dichloroethene
Concen: 20.464 ug/l
RT: 4.348 min Scan# 407
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40


Tgt Ion:	96	Resp:	37546
Ion	Ratio	Lower	Upper
96	100		
61	170.0	123.4	185.2
98	63.8	53.7	80.5








#11
Methyl Iodide
Concen: 19.608 ug/l
RT: 4.606 min Scan# 451
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020







Tgt Ion:142 Resp: 50473
Ion Ratio Lower Upper
142 100
127 50.5 23.2 69.5
141 10.8 7.2 21.7



#12
Methyl Acetate
Concen: 25.806 ug/l
RT: 5.042 min Scan# 525
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 43 Resp: 82925
Ion Ratio Lower Upper
43 100
74 22.8 20.5 30.7





#13

Acrolein

Concen: 120.806 ug/l

RT: 4.195 min Scan# 381

Delta R.T. 0.000 min

Lab File: VN078806.D

Acq: 08 Aug 2023 18:40

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC020

Tgt Ion: 56 Resp: 40428

Ion Ratio Lower Upper

56 100

55 70.5 59.0 88.4



#14

Acrylonitrile

Concen: 119.944 ug/l

RT: 5.730 min Scan# 642

Delta R.T. 0.000 min

Lab File: VN078806.D

Acq: 08 Aug 2023 18:40




Tgt Ion: 53 Resp: 110203


Ion Ratio Lower Upper

53 100

52 80.5 42.3 126.9

51 35.2 18.4 55.4





#15

Acetone

Concen: 108.176 ug/l

RT: 4.442 min Scan# 41

Delta R.T. 0.000 min

Lab File: VN078806.D

Acq: 08 Aug 2023 18:40

Instrument :

MSVOA_N

ClientSampleId :





VSTDICCC020

Tgt Ion: 58 Resp: 28047

Ion Ratio Lower Upper

58 100

43 328.0 227.5 341.3



#16

Carbon Disulfide

Concen: 21.237 ug/l

RT: 4.724 min Scan# 471

Delta R.T. 0.000 min

Lab File: VN078806.D




Acq: 08 Aug 2023 18:40


Tgt Ion: 76 Resp: 111045

Ion Ratio Lower Upper

76 100

78 7.5 7.7 11.5#









#17
 Allyl chloride
 Concen: 23.581 ug/l
 RT: 5.036 min Scan# 51
 Delta R.T. 0.000 min
 Lab File: VN078806.D
 Acq: 08 Aug 2023 18:40

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC020

Tgt Ion: 41 Resp: 60646




Ion	Ratio	Lower	Upper
41	100		
39	68.9	37.6	112.8
76	39.1	23.8	71.4




#18
 Methylene Chloride
 Concen: 21.186 ug/l
 RT: 5.283 min Scan# 566
 Delta R.T. 0.000 min
 Lab File: VN078806.D
 Acq: 08 Aug 2023 18:40

Tgt Ion: 84 Resp: 47269

Ion	Ratio	Lower	Upper
84	100		
49	109.1	88.1	132.1
51	39.4	24.0	36.0#
86	51.3	53.8	80.8#





#19

trans-1,2-Dichloroethene

Concen: 20.086 ug/l

RT: 5.795 min Scan# 61

Delta R.T. 0.000 min

Lab File: VN078806.D

Acq: 08 Aug 2023 18:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC020





Tgt Ion: 96 Resp: 42756

Ion Ratio Lower Upper

96 100

61 139.2 105.2 157.8

98 68.5 50.5 75.7



#20

Diisopropyl ether

Concen: 24.629 ug/l

RT: 6.683 min Scan# 804

Delta R.T. 0.000 min

Lab File: VN078806.D

Acq: 08 Aug 2023 18:40

Tgt Ion: 45 Resp: 146795




Ion Ratio Lower Upper


45 100

43 56.4 43.6 65.4

87 22.8 25.8 38.6#

59 13.4 10.6 16.0









#21
1,1-Dichloroethane
Concen: 21.850 ug/l
RT: 6.577 min Scan# 786
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:	63	Resp:	83556
Ion Ratio	Lower	Upper	
63	100		
98	5.6	2.7	8.1
100	5.6	2.5	7.4



#22
cis-1,2-Dichloroethene
Concen: 12.188 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion:	96	Resp:	26194
Ion Ratio	Lower	Upper	
96	100		
61	271.7	113.8	170.6#
98	77.5	52.2	78.4









#23
tert-Butyl Alcohol
Concen: 108.207 ug/l
RT: 5.536 min Scan# 609
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 59 Resp: 39298
Ion Ratio Lower Upper
59 100
52 0.0 0.1 0.1#



#24
Methyl tert-Butyl Ether
Concen: 20.428 ug/l
RT: 5.812 min Scan# 656
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 73 Resp: 126258
Ion Ratio Lower Upper
73 100
43 21.6 14.5 21.7









#25
Chloroform
Concen: 20.761 ug/l
RT: 7.977 min Scan# 1024
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 83 Resp: 82652
Ion Ratio Lower Upper
83 100
85 57.9 53.4 80.0



#26
Cyclohexane
Concen: 22.807 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 56 Resp: 63808
Ion Ratio Lower Upper
56 100
89 0.0 0.0 0.0
84 91.2 74.2 111.4









#27
1,2-Dichloroethane-d4
Concen: 32.252 ug/l
RT: 8.583 min Scan# 1127
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 65 Resp: 79883
Ion Ratio Lower Upper
65 100
67 52.9 42.9 64.3



#28
1,4-Difluorobenzene
Concen: 30.000 ug/l
RT: 9.106 min Scan# 1216
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 114 Resp: 185975
Ion Ratio Lower Upper
114 100
63 20.4 15.2 22.8
88 14.8 12.5 18.7









#29
1,1-Dichloropropene
Concen: 21.685 ug/l
RT: 8.383 min Scan# 1093
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 75 Resp: 61681
Ion Ratio Lower Upper
75 100
110 34.8 0.0 72.2
77 29.4 0.0 62.8



#30
2-Butanone
Concen: 125.723 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 43 Resp: 147386
Ion Ratio Lower Upper
43 100
57 8.1 6.8 10.2
72 24.2 23.8 35.6









#31
2,2-Dichloropropane
Concen: 19.927 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 77 Resp: 63260
Ion Ratio Lower Upper
77 100
97 22.1 18.8 28.2



#32
1,1,1-Trichloroethane
Concen: 19.784 ug/l
RT: 8.183 min Scan# 1059
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 97 Resp: 69151
Ion Ratio Lower Upper
97 100
99 66.4 51.3 76.9
61 49.7 34.6 52.0









#33
Carbon Tetrachloride
Concen: 19.973 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:117 Resp: 62426
Ion Ratio Lower Upper
117 100
119 98.6 81.4 122.0
121 28.8 23.6 35.4



#34
Benzene
Concen: 21.407 ug/l
RT: 8.612 min Scan# 1132
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 78 Resp: 182106
Ion Ratio Lower Upper
78 100
77 24.6 18.3 27.5
51 19.9 13.7 20.5









#35
Methacrylonitrile
Concen: 26.279 ug/l
RT: 7.789 min Scan# 992
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Tgt Ion: 41 Resp: 34461




Ion	Ratio	Lower	Upper
41	100		
39	64.4	33.9	101.6
67	69.0	45.9	137.6
52	27.8	18.2	54.6




#36
1,2-Dichloroethane
Concen: 21.497 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 62 Resp: 63887

Ion	Ratio	Lower	Upper
62	100		
98	9.7	7.4	11.0









#37
Trichloroethene
Concen: 19.029 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:130 Resp: 44449
Ion Ratio Lower Upper
130 100
95 103.9 78.2 117.4



#38
Methylcyclohexane
Concen: 20.334 ug/l
RT: 9.606 min Scan# 1301
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 83 Resp: 59023
Ion Ratio Lower Upper
83 100
55 74.5 34.0 102.0
98 46.6 23.0 69.0









#39
1,2-Dichloropropane
Concen: 21.313 ug/l
RT: 9.624 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 63 Resp: 49101
Ion Ratio Lower Upper
63 100
65 27.7 27.4 41.2



#40
Dibromomethane
Concen: 20.245 ug/l
RT: 9.712 min Scan# 1319
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 93 Resp: 31583
Ion Ratio Lower Upper
93 100
95 80.5 0.0 167.8
174 90.6 0.0 180.8









#41
Bromodichloromethane
Concen: 21.236 ug/l
RT: 9.894 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 83 Resp: 66664
Ion Ratio Lower Upper
83 100
85 66.0 54.1 81.1
127 8.2 7.8 11.6



#42
Vinyl Acetate
Concen: 124.787 ug/l
RT: 6.618 min Scan# 793
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 43 Resp: 469611
Ion Ratio Lower Upper
43 100
86 9.7 9.4 14.0









#43
Ethyl Acetate
Concen: 8.304 ug/l
RT: 7.565 min Scan# 91
Delta R.T. -0.012 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 43 Resp: 20493
Ion Ratio Lower Upper
43 100
45 41.3 10.9 16.3#



#44
Isopropyl Acetate
Concen: 25.258 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 43 Resp: 107010
Ion Ratio Lower Upper
43 100
61 10.7 23.0 34.6#









#45
1,4-Dioxane
Concen: 422.840 ug/l
RT: 9.700 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 88 Resp: 15982
Ion Ratio Lower Upper
88 100
43 35.4 19.0 28.6#
58 73.9 48.2 72.2#



#46
Methyl methacrylate
Concen: 23.964 ug/l
RT: 9.688 min Scan# 1315
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 41 Resp: 46397
Ion Ratio Lower Upper
41 100
69 90.4 49.0 147.0
39 54.2 31.6 94.8









#47
n-amyl acetate
Concen: 24.305 ug/l
RT: 12.500 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 43 Resp: 79278
Ion Ratio Lower Upper
43 100
55 24.1 21.5 32.3



#48
t-1,3-Dichloropropene
Concen: 20.289 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 75 Resp: 66659
Ion Ratio Lower Upper
75 100
77 30.2 26.1 39.1









#49
 cis-1,3-Dichloropropene
 Concen: 20.938 ug/l
 RT: 10.318 min Scan# 1422
 Delta R.T. 0.000 min
 Lab File: VN078806.D
 Acq: 08 Aug 2023 18:40




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC020


Tgt Ion:	75	Resp:	74589
Ion	Ratio	Lower	Upper
75	100		
77	30.4	24.0	36.0
39	54.2	35.1	52.7#



#50
 1,1,2-Trichloroethane
 Concen: 20.349 ug/l
 RT: 11.024 min Scan# 1542
 Delta R.T. 0.000 min
 Lab File: VN078806.D
 Acq: 08 Aug 2023 18:40

Tgt Ion:	97	Resp:	44160
Ion	Ratio	Lower	Upper
97	100		
83	80.1	66.8	100.2
85	55.1	42.5	63.7
99	56.9	48.3	72.5









#51
Ethyl methacrylate
Concen: 20.289 ug/l
RT: 10.883 min Scan# 1518
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 69 Resp: 60595
Ion Ratio Lower Upper
69 100
41 81.5 26.0 78.0#
39 45.6 16.3 48.9



#52
1,3-Dichloropropane
Concen: 19.226 ug/l
RT: 11.171 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 76 Resp: 75854
Ion Ratio Lower Upper
76 100
78 32.3 0.0 66.2









#53
Dibromochloromethane
Concen: 18.612 ug/l
RT: 11.365 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:129 Resp: 44805
Ion Ratio Lower Upper
129 100
127 81.7 39.4 118.1



#54
1,2-Dibromoethane
Concen: 19.726 ug/l
RT: 11.477 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion:107 Resp: 42686
Ion Ratio Lower Upper
107 100
109 90.9 74.6 112.0









#55
2-Chloroethyl vinyl ether
Concen: 114.352 ug/l
RT: 10.165 min Scan# 112133
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:	63	Resp:	112133
Ion	Ratio	Lower	Upper
63	100		
65	31.9	26.3	39.5
106	27.2	23.8	35.6



#56
Bromoform
Concen: 17.988 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion:	173	Resp:	28780
Ion	Ratio	Lower	Upper
173	100		
175	48.9	40.7	61.1
254	8.6	6.5	9.7









#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 11.871 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:117 Resp: 156856
Ion Ratio Lower Upper
117 100
82 59.3 43.9 65.9
119 34.5 26.2 39.4



#58
4-Methyl-2-Pentanone
Concen: 123.975 ug/l
RT: 10.453 min Scan# 1445
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 43 Resp: 315458
Ion Ratio Lower Upper
43 100
58 35.5 32.6 48.8
85 15.4 16.2 24.2#









#59
2-Hexanone
Concen: 117.699 ug/l
RT: 11.200 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 43 Resp: 221370
Ion Ratio Lower Upper
43 100
58 50.6 45.9 68.9
57 16.5 15.8 23.6



#60
4-Bromofluorobenzene
Concen: 29.197 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 95 Resp: 69391
Ion Ratio Lower Upper
95 100
174 77.0 61.7 92.5
176 74.2 58.1 87.1









#61
Tetrachloroethene
Concen: 17.910 ug/l
RT: 11.112 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:164 Resp: 36710
Ion Ratio Lower Upper
164 100
166 122.5 99.1 148.7
129 97.5 85.6 128.4
131 98.0 77.6 116.4



#62
Toluene
Concen: 20.064 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 91 Resp: 192478
Ion Ratio Lower Upper
91 100
92 56.6 45.2 67.8









#63
Toluene-d8
Concen: 30.203 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 98 Resp: 237195
Ion Ratio Lower Upper
98 100
100 64.4 49.7 74.5



#64
Chlorobenzene
Concen: 20.010 ug/l
RT: 11.900 min Scan# 1691
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 112 Resp: 111495
Ion Ratio Lower Upper
112 100
114 30.1 25.9 38.9









#65
1,1,1,2-Tetrachloroethane
Concen: 19.736 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:131 Resp: 43056
Ion Ratio Lower Upper
131 100
133 96.3 0.0 191.8
119 64.8 0.0 132.0



#66
Ethyl Benzene
Concen: 21.081 ug/l
RT: 11.971 min Scan# 1703
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 91 Resp: 204995
Ion Ratio Lower Upper
91 100
106 27.7 25.6 38.4









#67
m/p-Xylenes
Concen: 39.958 ug/l
RT: 12.077 min Scan# 1721
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:106 Resp: 146780
Ion Ratio Lower Upper
106 100
91 201.1 162.5 243.7



#68
o-Xylene
Concen: 19.732 ug/l
RT: 12.406 min Scan# 1777
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion:106 Resp: 74229
Ion Ratio Lower Upper
106 100
91 208.4 106.1 318.3









#69
Styrene
Concen: 18.611 ug/l
RT: 12.418 min Scan# 1779
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:104 Resp: 111990
Ion Ratio Lower Upper
104 100
78 52.2 40.1 60.1
103 59.0 45.1 67.7



#70
Isopropylbenzene
Concen: 19.479 ug/l
RT: 12.700 min Scan# 1827
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion:105 Resp: 180232
Ion Ratio Lower Upper
105 100
120 27.0 18.3 34.1









#71
1,1,2,2-Tetrachloroethane
Concen: 20.305 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 83 Resp: 61793
Ion Ratio Lower Upper
83 100
131 9.6 4.8 14.4
85 64.0 32.6 98.0



#72
1,2,3-Trichloropropane
Concen: 29.879 ug/l
RT: 13.000 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 75 Resp: 74030
Ion Ratio Lower Upper
75 100
77 124.0 0.0 387.4









#73
Bromobenzene
Concen: 18.949 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:156 Resp: 42693
Ion Ratio Lower Upper
156 100
77 215.0 0.0 420.8
158 94.8 0.0 193.2



#74
n-propylbenzene
Concen: 19.793 ug/l
RT: 13.041 min Scan# 1885
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 91 Resp: 204759
Ion Ratio Lower Upper
91 100
120 22.8 0.0 44.6









#75
2-Chlorotoluene
Concen: 17.924 ug/l
RT: 13.129 min Scan# 1900
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion: 91 Resp: 127452
Ion Ratio Lower Upper
91 100
126 32.0 0.0 67.2



#76
1,3,5-Trimethylbenzene
Concen: 18.013 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion: 105 Resp: 148955
Ion Ratio Lower Upper
105 100
120 46.5 0.0 97.2









#77
t-1,4-Dichloro-2-butene
Concen: 19.427 ug/l
RT: 12.741 min Scan# 1834
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:	75	Resp:	17182
Ion	Ratio	Lower	Upper
75	100		
53	112.2	41.6	124.9
89	46.1	25.1	75.4



#78
4-Chlorotoluene
Concen: 17.569 ug/l
RT: 13.229 min Scan# 1917
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion:	91	Resp:	119075
Ion	Ratio	Lower	Upper
91	100		
126	32.5	0.0	68.8









#79
tert-butylbenzene
Concen: 17.385 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:119 Resp: 123189
Ion Ratio Lower Upper
119 100
91 63.9 0.0 126.6
134 24.2 0.0 44.0



#80
1,2,4-Trimethylbenzene
Concen: 17.861 ug/l
RT: 13.488 min Scan# 1961
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion:105 Resp: 140130
Ion Ratio Lower Upper
105 100
120 43.3 0.0 88.2









#81
 sec-Butylbenzene
 Concen: 17.886 ug/l
 RT: 13.624 min Scan# 1984
 Delta R.T. 0.000 min
 Lab File: VN078806.D
 Acq: 08 Aug 2023 18:40




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC020


Tgt Ion:105 Resp: 164005
 Ion Ratio Lower Upper
 105 100
 134 19.2 0.0 38.6



#82
 p-Isopropyltoluene
 Concen: 17.289 ug/l
 RT: 13.735 min Scan# 2003
 Delta R.T. 0.000 min
 Lab File: VN078806.D
 Acq: 08 Aug 2023 18:40

Tgt Ion:119 Resp: 127838
 Ion Ratio Lower Upper
 119 100
 134 25.8 0.0 52.4
 91 23.6 0.0 48.2









#83
 1,3-Dichlorobenzene
 Concen: 17.601 ug/l
 RT: 13.735 min Scan# 2003
 Delta R.T. 0.000 min
 Lab File: VN078806.D
 Acq: 08 Aug 2023 18:40




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC020


Tgt Ion:146 Resp: 69911
 Ion Ratio Lower Upper
 146 100
 111 41.9 21.6 64.6
 148 64.5 31.9 95.5



#84
 1,4-Dichlorobenzene
 Concen: 17.671 ug/l
 RT: 13.818 min Scan# 2017
 Delta R.T. 0.000 min
 Lab File: VN078806.D
 Acq: 08 Aug 2023 18:40

Tgt Ion:146 Resp: 65844
 Ion Ratio Lower Upper
 146 100
 111 38.5 19.4 58.2
 148 61.7 31.4 94.3









#85
n-Butylbenzene
Concen: 17.978 ug/l
RT: 14.059 min Scan# 2058
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:	91	Resp:	92573
Ion	Ratio	Lower	Upper
91	100		
92	51.5	0.0	99.6
134	25.3	0.0	50.0



#86
Hexachloroethane
Concen: 19.046 ug/l
RT: 14.341 min Scan# 2106
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion:	117	Resp:	26063
Ion	Ratio	Lower	Upper
117	100		
201	60.5	34.2	102.6





#87

1,2-Dichlorobenzene

Concen: 18.383 ug/l

RT: 14.112 min Scan# 2067

Delta R.T. 0.000 min

Lab File: VN078806.D

Acq: 08 Aug 2023 18:40

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC020





Tgt Ion:146 Resp: 68778

Ion Ratio Lower Upper

146 100

111 40.6 22.4 67.3

148 64.6 33.4 100.1



#88

1,2-Dibromo-3-Chloropropane

Concen: 17.075 ug/l

RT: 14.723 min Scan# 2171

Delta R.T. 0.000 min

Lab File: VN078806.D

Acq: 08 Aug 2023 18:40




Tgt Ion: 75 Resp: 8617


Ion Ratio Lower Upper

75 100

155 88.4 68.2 102.2

157 107.5 84.7 127.1









#89
1,2,4-Trichlorobenzene
Concen: 17.981 ug/l
RT: 15.400 min Scan# 21
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020


Tgt Ion:180 Resp: 19027
Ion Ratio Lower Upper
180 100
182 93.0 73.7 110.5
145 35.8 25.0 37.6



#90
Hexachlorobutadiene
Concen: 17.647 ug/l
RT: 15.506 min Scan# 2304
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion:225 Resp: 14045
Ion Ratio Lower Upper
225 100
223 60.2 0.0 130.0
227 64.5 0.0 124.0









#91
Naphthalene
Concen: 18.200 ug/l
RT: 15.647 min Scan# 2328
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC020

Tgt Ion:	128	Resp:	58324
Ion	Ratio	Lower	Upper
128	100		
127	14.2	10.2	15.2
129	10.6	9.0	13.4



#92
1,2,3-Trichlorobenzene
Concen: 18.530 ug/l
RT: 15.847 min Scan# 2362
Delta R.T. 0.000 min
Lab File: VN078806.D
Acq: 08 Aug 2023 18:40

Tgt Ion:	180	Resp:	22358
Ion	Ratio	Lower	Upper
180	100		
182	89.6	0.0	196.4
145	14.4	0.0	71.0

