

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080823\
 Data File : VN078792.D
 Acq On : 08 Aug 2023 10:39
 Operator : JC\MD
 Sample : VN0808MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0808MBL01

Quant Time: Aug 09 02:47:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080723W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 08 05:03:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	120562	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	225094	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	189947	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	47269	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	91521	50.637	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.280%
35) Dibromofluoromethane	8.177	113	76549	50.358	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.720%
50) Toluene-d8	10.571	98	273836	48.953	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.900%
62) 4-Bromofluorobenzene	12.853	95	68516	40.984	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	81.960%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080823\
Data File : VN078792.D
Acq On : 08 Aug 2023 10:39
Operator : JC\MD
Sample : VN0808MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0808MBL01

Quant Time: Aug 09 02:47:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080723W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 08 05:03:52 2023
Response via : Initial Calibration

