

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080823\
 Data File : VN078809.D
 Acq On : 08 Aug 2023 19:52
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Aug 09 04:42:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 09 04:37:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	28793	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.112	114	157449	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	123985	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.588	65	70517	31.887	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	106.300%
60) 4-Bromofluorobenzene	12.853	95	38457	20.471	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	68.233%
63) Toluene-d8	10.571	98	187476	30.201	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.667%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080823\
Data File : VN078809.D
Acq On : 08 Aug 2023 19:52
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Aug 09 04:42:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 09 04:37:34 2023
Response via : Initial Calibration

