

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080824\
 Data File : VN083149.D
 Acq On : 08 Aug 2024 12:33
 Operator : JC\MD
 Sample : P3415-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MLS-15-106-121

Quant Time: Aug 08 13:33:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	144631	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	272402	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	269888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	120407	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	113701	55.231	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.460%
35) Dibromofluoromethane	8.165	113	91670	53.915	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.820%
50) Toluene-d8	10.565	98	348640	54.969	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.940%
62) 4-Bromofluorobenzene	12.847	95	144668	58.507	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	117.020%
Target Compounds						
52) Toluene	10.630	92	21285	4.397	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080824\
Data File : VN083149.D
Acq On : 08 Aug 2024 12:33
Operator : JC\MD
Sample : P3415-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MLS-15-106-121

Quant Time: Aug 08 13:33:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

