

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080824\
 Data File : VN083152.D
 Acq On : 08 Aug 2024 13:46
 Operator : JC\MD
 Sample : P3415-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MLS-15-9999

Quant Time: Aug 08 14:25:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	142360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	272221	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	279265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	123786	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	114883	56.695	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 113.400%	
35) Dibromofluoromethane	8.171	113	93314	54.918	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.840%	
50) Toluene-d8	10.565	98	356037	56.173	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 112.340%	
62) 4-Bromofluorobenzene	12.847	95	147813	59.819	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 119.640%	
Target Compounds						
52) Toluene	10.635	92	20462	4.229	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080824\
Data File : VN083152.D
Acq On : 08 Aug 2024 13:46
Operator : JC\MD
Sample : P3415-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MLS-15-9999

Quant Time: Aug 08 14:25:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

