

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080824\
 Data File : VN083157.D
 Acq On : 08 Aug 2024 15:46
 Operator : JC\MD
 Sample : P3393-17ME
 Misc : 7.71g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-849-GI-SO-10-10.5-072624ME

Quant Time: Aug 09 01:04:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	135153	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	256173	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	261585	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	126534	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	104463	54.302	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.600%	
35) Dibromofluoromethane	8.165	113	86476	54.082	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.160%	
50) Toluene-d8	10.565	98	331999	55.662	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.320%	
62) 4-Bromofluorobenzene	12.847	95	142064	61.094	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 122.180%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.336	101	283337	194.295	ug/l	99
27) cis-1,2-Dichloroethene	7.483	96	8126	4.349	ug/l	82
44) Trichloroethene	9.353	130	5923638	3453.791	ug/l	90
64) Tetrachloroethene	11.106	164	43133	24.898	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080824\
Data File : VN083157.D
Acq On : 08 Aug 2024 15:46
Operator : JC\MD
Sample : P3393-17ME
Misc : 7.71g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-849-GI-SO-10-10.5-072624ME

Quant Time: Aug 09 01:04:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

