

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080824\
 Data File : VN083158.D
 Acq On : 08 Aug 2024 16:10
 Operator : JC\MD
 Sample : P3393-18ME
 Misc : 6.89g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-849-GI-SO-12.5-13-072624ME

Quant Time: Aug 09 01:05:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	136109	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.095	114	254132	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	261069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	125130	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	106337	54.888	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.780%	
35) Dibromofluoromethane	8.159	113	85995	54.213	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.420%	
50) Toluene-d8	10.565	98	331534	56.030	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 112.060%	
62) 4-Bromofluorobenzene	12.847	95	143120	62.043	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 124.080%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.330	101	107289	73.056	ug/l	99
16) Acetone	4.460	43	6760	8.917	ug/l	91
27) cis-1,2-Dichloroethene	7.489	96	8374	4.450	ug/l	83
44) Trichloroethene	9.353	130	6633474	3898.725	ug/l	91
64) Tetrachloroethene	11.100	164	29715	17.186	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080824\
Data File : VN083158.D
Acq On : 08 Aug 2024 16:10
Operator : JC\MD
Sample : P3393-18ME
Misc : 6.89g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-849-GI-SO-12.5-13-072624ME

Quant Time: Aug 09 01:05:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

