

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080824\
 Data File : VN083160.D
 Acq On : 08 Aug 2024 16:58
 Operator : JC\MD
 Sample : P3393-20
 Misc : 7.93g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-849-GI-SO-13.5-14-072624

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
 Supervised By :Semsettin Yesilyurt 08/09/2024

Quant Time: Aug 09 01:06:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	139550	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	267935	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	273286	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	135972	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	112887	56.832	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 113.660%			
35) Dibromofluoromethane	8.165	113	89296	53.394	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.780%			
50) Toluene-d8	10.565	98	341524	54.745	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 109.500%			
62) 4-Bromofluorobenzene	12.847	95	148290	60.972	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 121.940%			
Target Compounds					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.330	101	1069344	710.186	ug/l	99
16) Acetone	4.465	43	5787	7.446	ug/l	90
27) cis-1,2-Dichloroethene	7.483	96	19767	10.246	ug/l	82
30) Chloroform	7.959	83	5078	1.632	ug/l	98
38) Carbon Tetrachloride	8.353	117	8847	3.105	ug/l	89
44) Trichloroethene	9.347	130	35092174m	19562.383	ug/l	
55) 1,1,2-Trichloroethane	11.018	97	8614	5.048	ug/l	96
64) Tetrachloroethene	11.106	164	649072	358.622	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080824\
Data File : VN083160.D
Acq On : 08 Aug 2024 16:58
Operator : JC\MD
Sample : P3393-20
Misc : 7.93g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-849-GI-SO-13.5-14-072624

Quant Time: Aug 09 01:06:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/09/2024
Supervised By :Semsettin Yesilyurt 08/09/2024

