

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080824\
 Data File : VN083161.D
 Acq On : 08 Aug 2024 17:22
 Operator : JC\MD
 Sample : P3393-21
 Misc : 8.38g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-850-GI-SO-10-10.5-072624

Quant Time: Aug 09 01:07:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	141514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	265572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	269257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	128814	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	109529	54.376	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.760%	
35) Dibromofluoromethane	8.165	113	88090	53.142	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.280%	
50) Toluene-d8	10.565	98	346098	55.972	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.940%	
62) 4-Bromofluorobenzene	12.847	95	146897	60.937	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 121.880%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.324	101	3146	2.060	ug/l	96
27) cis-1,2-Dichloroethene	7.489	96	2915	1.490	ug/l	88
44) Trichloroethene	9.347	130	246049	138.382	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080824\
Data File : VN083161.D
Acq On : 08 Aug 2024 17:22
Operator : JC\MD
Sample : P3393-21
Misc : 8.38g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-850-GI-SO-10-10.5-072624

Quant Time: Aug 09 01:07:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

