

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN080918\
 Data File : VN050479.D
 Acq On : 9 Aug 2018 12:33
 Operator : MD\SY
 Sample : J4310-03
 Misc : 3.48g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 C-2

Quant Time: Aug 10 04:36:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	721461	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1104739	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1010839	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	464503	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	460764	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
35) Dibromofluoromethane	7.59	113	414333	46.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.48%	
50) Toluene-d8	10.09	98	1545973	46.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.78%	
62) 4-Bromofluorobenzene	12.40	95	502805	44.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.52%	
Target Compounds						
10) Methyl Iodide	3.94	142	6286	1.72	ug/l #	95
18) Methyl Acetate	4.33	43	8915	1.00	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080918\
Data File : VN050479.D
Acq On : 9 Aug 2018 12:33
Operator : MD\SY
Sample : J4310-03
Misc : 3.48g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
C-2

Quant Time: Aug 10 04:36:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

