

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080918\
 Data File : VN050503.D
 Acq On : 9 Aug 2018 23:17
 Operator : MD\SY
 Sample : J4352-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIPBLANK

Manual Integrations
 APPROVED

MMDadoda
 8/10/2018 10:53:17 AM

Quant Time: Aug 10 08:55:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N072518W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jul 26 15:21:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	127276	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	652745	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	545992	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	258788	29.91	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	99.70%
60) 4-Bromofluorobenzene	12.40	95	195062	23.25	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	77.50%
63) Toluene-d8	10.09	98	756554	29.01	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	96.70%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080918\
Data File : VN050503.D
Acq On : 9 Aug 2018 23:17
Operator : MD\SY
Sample : J4352-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIPBLANK

Manual Integrations
APPROVED

MMDadoda
8/10/2018 10:53:17 AM

Quant Time: Aug 10 08:55:44 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N072518W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jul 26 15:21:50 2018
Response via : Initial Calibration

