

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080919\
 Data File : VN057142.D
 Acq On : 9 Aug 2019 13:24
 Operator : JC/SP
 Sample : K4281-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Aug 10 07:39:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1418151	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1973338	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1621502	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	385350	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.44	65	650338	44.59	ug/l	0.00
Spiked Amount			Recovery	=	89.18%	
35) Dibromofluoromethane	6.99	113	596902	48.20	ug/l	0.00
Spiked Amount			Recovery	=	96.40%	
50) Toluene-d8	9.56	98	2288253	48.30	ug/l	0.00
Spiked Amount			Recovery	=	96.60%	
62) 4-Bromofluorobenzene	11.92	95	577382	37.82	ug/l	0.00
Spiked Amount			Recovery	=	75.64%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080919\
Data File : VN057142.D
Acq On : 9 Aug 2019 13:24
Operator : JC/SP
Sample : K4281-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Aug 10 07:39:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

