

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080919\
 Data File : VN057150.D
 Acq On : 9 Aug 2019 16:38
 Operator : JC/SP
 Sample : K4262-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-19D_080819

Quant Time: Aug 10 07:59:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1369616	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1912944	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1561749	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	375179	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	629585	44.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.40%	
35) Dibromofluoromethane	6.99	113	568415	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
50) Toluene-d8	9.56	98	2206681	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
62) 4-Bromofluorobenzene	11.92	95	567831	38.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.74%	
Target Compounds						
16) Acetone	3.48	43	24209	5.691	ug/l	95
18) Methyl Acetate	3.90	43	16311	0.638	ug/l	99
44) Trichloroethene	8.27	130	58232	4.169	ug/l	97
64) Tetrachloroethene	10.11	164	2046365	157.556	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080919\
Data File : VN057150.D
Acq On : 9 Aug 2019 16:38
Operator : JC/SP
Sample : K4262-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-19D_080819

Quant Time: Aug 10 07:59:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

