

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN080919\
 Data File : VN057156.D
 Acq On : 9 Aug 2019 19:11
 Operator : JC/SP
 Sample : K4290-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WP-1

Quant Time: Aug 10 08:13:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:49:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	1304531	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.01	114	1801737	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.91	117	1475736	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.88	152	348536	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	606240	45.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.38%	
35) Dibromofluoromethane	6.99	113	543427	48.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.12%	
50) Toluene-d8	9.56	98	2086852	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
62) 4-Bromofluorobenzene	11.92	95	529810	38.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.02%	
Target Compounds						
16) Acetone	3.48	43	14418	3.559	ug/l	99
18) Methyl Acetate	3.90	43	19093	0.937	ug/l #	76
64) Tetrachloroethene	10.11	164	43723	3.563	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN080919\
Data File : VN057156.D
Acq On : 9 Aug 2019 19:11
Operator : JC/SP
Sample : K4290-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WP-1

Quant Time: Aug 10 08:13:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080719W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:49:25 2019
Response via : Initial Calibration

