

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
 Data File : VN068035.D
 Acq On : 9 Aug 2021 12:56
 Operator : JC/MD
 Sample : VN0809WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0809WBL01

Quant Time: Aug 10 03:48:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	334296	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	632264	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	618794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	220465	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	253746	53.617	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 107.240%		
35) Dibromofluoromethane	8.024	113	197853	53.695	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 107.400%		
50) Toluene-d8	10.446	98	804136	47.497	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 95.000%		
62) 4-Bromofluorobenzene	12.733	95	259668	42.542	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 85.080%		
Target Compounds						
						Qvalue
95) Naphthalene	15.509	128	12404	4.385	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
Data File : VN068035.D
Acq On : 9 Aug 2021 12:56
Operator : JC/MD
Sample : VN0809WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0809WBL01

Quant Time: Aug 10 03:48:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

