

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
 Data File : VN068038.D
 Acq On : 9 Aug 2021 15:06
 Operator : JC/MD
 Sample : PB138251TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138251TB

Manual Integrations
 APPROVED

MMDadoda
 8/10/2021 7:28:16 PM

Quant Time: Aug 10 03:50:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	335762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	627660	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	660884	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	234364	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	251934	53.002	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.000%
35) Dibromofluoromethane	8.024	113	195497	53.445	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.900%
50) Toluene-d8	10.443	98	831968	49.338	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.680%
62) 4-Bromofluorobenzene	12.734	95	281781	46.040	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.080%
Target Compounds						
						Qvalue
16) Acetone	4.291	43	34651	19.979	ug/l	93
20) Methylene Chloride	5.093	84	42877	10.482	ug/l	86
25) 2-Butanone	7.354	43	25017m	9.942	ug/l	
43) Isopropyl Acetate	8.566	43	43616	4.790	ug/l #	92
95) Naphthalene	15.509	128	27204	5.730	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
Data File : VN068038.D
Acq On : 9 Aug 2021 15:06
Operator : JC/MD
Sample : PB138251TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138251TB

Manual Integrations
APPROVED

MMDadoda
8/10/2021 7:28:16 PM

Quant Time: Aug 10 03:50:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

