

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
 Data File : VN068040.D
 Acq On : 9 Aug 2021 15:55
 Operator : JC/MD
 Sample : PB138251ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138251ZHE#02

Manual Integrations
 APPROVED

MMDadoda
 8/10/2021 7:28:18 PM

Quant Time: Aug 10 03:51:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	326827	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	643471	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	659053	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	229930	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	251340	54.323	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 108.640%		
35) Dibromofluoromethane	8.024	113	190856	50.894	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 101.780%		
50) Toluene-d8	10.443	98	830501	48.143	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 96.280%		
62) 4-Bromofluorobenzene	12.733	95	274723	44.028	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 88.060%		
Target Compounds						Qvalue
16) Acetone	4.296	43	33134	19.627	ug/l	94
18) Methyl Acetate	4.870	43	4778m	1.077	ug/l	
20) Methylene Chloride	5.095	84	41104	10.319	ug/l	89
25) 2-Butanone	7.345	43	23537	9.610	ug/l	98
43) Isopropyl Acetate	8.566	43	42645	4.569	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
Data File : VN068040.D
Acq On : 9 Aug 2021 15:55
Operator : JC/MD
Sample : PB138251ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138251ZHE#02

Manual Integrations
APPROVED

MMDadoda
8/10/2021 7:28:18 PM

Quant Time: Aug 10 03:51:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

