

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
 Data File : VN068041.D
 Acq On : 9 Aug 2021 16:20
 Operator : JC/MD
 Sample : PB138251ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB138251ZHE#03

Manual Integrations
 APPROVED

MMDadoda
 8/10/2021 7:28:20 PM

Quant Time: Aug 10 03:51:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	315764	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	619145	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	644657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	222484	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	244617	54.722	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	= 109.440%		
35) Dibromofluoromethane	8.027	113	190122	52.690	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	= 105.380%		
50) Toluene-d8	10.443	98	802646	48.339	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	= 96.680%		
62) 4-Bromofluorobenzene	12.736	95	267761	44.533	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	= 89.060%		
Target Compounds						
						Qvalue
16) Acetone	4.296	43	30344	18.604	ug/l	94
18) Methyl Acetate	4.867	43	5199m	1.226	ug/l	
20) Methylene Chloride	5.095	84	35623	9.227	ug/l	88
25) 2-Butanone	7.343	43	24384m	10.305	ug/l	
43) Isopropyl Acetate	8.568	43	43848	4.882	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
Data File : VN068041.D
Acq On : 9 Aug 2021 16:20
Operator : JC/MD
Sample : PB138251ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB138251ZHE#03

Manual Integrations
APPROVED

MMDadoda
8/10/2021 7:28:20 PM

Quant Time: Aug 10 03:51:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

