

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
 Data File : VN068045.D
 Acq On : 9 Aug 2021 17:58
 Operator : JC/MD
 Sample : M3318-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SPLP-AMENDED-COM-9

Manual Integrations
 APPROVED

MMDadoda
 8/10/2021 7:28:26 PM

Quant Time: Aug 10 03:53:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	292226	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	570383	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	546049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	185162	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	229150	55.391	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 110.780%			
35) Dibromofluoromethane	8.024	113	160406	48.255	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 96.520%			
50) Toluene-d8	10.443	98	717258	47.005	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 94.020%			
62) 4-Bromofluorobenzene	12.733	95	230141	41.882	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 83.760%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	31996	21.197	ug/l	94
18) Methyl Acetate	4.859	43	7137m	1.868	ug/l	
20) Methylene Chloride	5.093	84	12401	3.292	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
Data File : VN068045.D
Acq On : 9 Aug 2021 17:58
Operator : JC/MD
Sample : M3318-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SPLP-AMENDED-COM-9

Manual Integrations
APPROVED

MMDadoda
8/10/2021 7:28:26 PM

Quant Time: Aug 10 03:53:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

