

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
 Data File : VN068047.D
 Acq On : 9 Aug 2021 18:47
 Operator : JC/MD
 Sample : M3308-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11930

Quant Time: Aug 10 03:53:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	291922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	568971	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	597125	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	214257	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	230165	55.694	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.380%
35) Dibromofluoromethane	8.027	113	173088	52.200	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.400%
50) Toluene-d8	10.443	98	747586	48.941	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.880%
62) 4-Bromofluorobenzene	12.734	95	258797	46.581	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.160%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.390	59	8464	15.286	ug/l #	94
16) Acetone	4.291	43	537597	356.521	ug/l	92
18) Methyl Acetate	4.865	43	13944	3.751	ug/l	94
20) Methylene Chloride	5.093	84	38571	10.855	ug/l	92
25) 2-Butanone	7.343	43	161037	73.612	ug/l	94
43) Isopropyl Acetate	8.566	43	31448	3.810	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
Data File : VN068047.D
Acq On : 9 Aug 2021 18:47
Operator : JC/MD
Sample : M3308-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11930

Quant Time: Aug 10 03:53:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

