

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
 Data File : VN068034.D
 Acq On : 9 Aug 2021 11:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 10 03:34:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	50.000	45.881	8.2	98	0.00
3 P	Chloromethane	50.000	45.216	9.6	99	0.00
4 C	Vinyl Chloride	50.000	41.618	16.8#	89	0.00
5 T	Bromomethane	50.000	46.881	6.2	115	-0.01
6 T	Chloroethane	50.000	40.533	18.9	89	0.00
7 T	Trichlorofluoromethane	50.000	39.557	20.9	88	0.00
8 T	Diethyl Ether	50.000	51.979	-4.0	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	45.272	9.5	100	0.00
10 T	Methyl Iodide	50.000	48.640	2.7	109	0.00
11 T	Tert butyl alcohol	250.000	274.850	-9.9	116	0.00
12 CM	1,1-Dichloroethene	50.000	46.244	7.5#	99	0.00
13 T	Acrolein	250.000	301.659	-20.7	119	0.00
14 T	Allyl chloride	50.000	49.807	0.4	107	0.00
15 T	Acrylonitrile	250.000	283.455	-13.4	117	0.00
16 T	Acetone	250.000	246.095	1.6	120	0.00
17 T	Carbon Disulfide	50.000	43.639	12.7	95	0.00
18 T	Methyl Acetate	50.000	55.071	-10.1	121	0.00
19 T	Methyl tert-butyl Ether	50.000	55.683	-11.4	114	0.00
20						

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
 Data File : VN068034.D
 Acq On : 9 Aug 2021 11:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 10 03:34:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 29 15:02:35 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1180.121	-18.0	119	0.00
50 S	Toluene-d8	50.000	49.317	1.4	118	0.00
51 T	4-Methyl-2-Pentanone	250.000	248.783	0.5	115	0.00
52 CM	Toluene	50.000	54.891	-9.8#	107	0.00
53 T	t-1,3-Dichloropropene	50.000	50.142	-0.3	116	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.432	1.1	112	0.00
55 T	1,1,2-Trichloroethane	50.000	57.913	-15.8	117	0.00
56 T	Ethyl methacrylate	50.000	50.181	-0.4	117	0.00
57 T	1,3-Dichloropropane	50.000	57.027	-14.1	113	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	204.609	18.2	96	0.00
59 T	2-Hexanone	250.000	251.182	-0.5	117	0.00
60 T	Dibromochloromethane	50.000	50.221	-0.4	119	0.00
61 T	1,2-Dibromoethane	50.000	58.676	-17.4	115	0.00
62 S	4-Bromofluorobenzene	50.000	51.311	-2.6	126	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	116	0.00
64 T	Tetrachloroethene	50.000	45.700	8.6	101	0.00
65 PM	Chlorobenzene	50.000	51.937	-3.9	115	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.420	1.2	117	0.00
67 C	Ethyl Benzene	50.000	51.060	-2.1#	108	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
Data File : VN068034.D
Acq On : 9 Aug 2021 11:57
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 10 03:34:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 29 15:02:35 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.001	-0.0	128	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6