


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080921\
Data File : VN068033.D
Acq On : 9 Aug 2021 9:07
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072921W.M
Title : SW846 8260
Last Update : Thu Jul 29 15:02:35 2021



AutoFind: Scans 4005, 4006, 4007; Background Corrected with Scan 3989

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.3	16046	PASS
75	95	30	60	56.2	40496	PASS
95	95	100	100	100.0	72115	PASS
96	95	5	9	6.5	4693	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	72.7	52395	PASS
175	174	5	9	8.5	4479	PASS
176	174	95	101	96.6	50616	PASS
177	176	5	9	7.8	3948	PASS