

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080922\
 Data File : VN073784.D
 Acq On : 09 Aug 2022 12:58
 Operator : JC\MD
 Sample : PB146833ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146833ZHE#01

Quant Time: Aug 09 21:07:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 29 02:20:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	393207	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	645893	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	640696	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	296092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	226981	47.368	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.740%
35) Dibromofluoromethane	7.957	113	217123	53.534	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.060%
50) Toluene-d8	10.380	98	739452	46.331	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.660%
62) 4-Bromofluorobenzene	12.674	95	277883	51.761	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.520%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	117653	45.713	ug/l	100
18) Methyl Acetate	4.792	43	8314	1.298	ug/l #	88
20) Methylene Chloride	5.010	84	30238	5.292	ug/l	98
25) 2-Butanone	7.269	43	24288	6.077	ug/l	95
37) Ethyl Acetate	7.345	43	69420	9.119	ug/l	98
43) Isopropyl Acetate	8.498	43	264650	25.738	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080922\
Data File : VN073784.D
Acq On : 09 Aug 2022 12:58
Operator : JC\MD
Sample : PB146833ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146833ZHE#01

Quant Time: Aug 09 21:07:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 29 02:20:25 2022
Response via : Initial Calibration

