

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080923\
 Data File : VN078820.D
 Acq On : 09 Aug 2023 15:08
 Operator : JC\MD
 Sample : 03940-02DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 002-35TH-AVE(JUNE)DL

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/10/2023
 Supervised By : Mahesh Dadoda 08/10/2023

Quant Time: Aug 10 05:29:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 09 05:05:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	32355	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.112	114	163241	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	135504	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.588	65	76680	29.437	ug/l	0.00
Spiked Amount 30.000	Range	91 - 110	Recovery	=	98.133%	
60) 4-Bromofluorobenzene	12.859	95	41255	21.690	ug/l	0.00
Spiked Amount 30.000	Range	63 - 112	Recovery	=	72.300%	
63) Toluene-d8	10.576	98	197043	29.675	ug/l	0.00
Spiked Amount 30.000	Range	91 - 112	Recovery	=	98.900%	
Target Compounds						
15) Acetone	4.447	58	20724	77.941	ug/l	Qvalue 99
30) 2-Butanone	7.500	43	5968m	4.591	ug/l	
62) Toluene	10.635	91	262583	32.915	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080923\
Data File : VN078820.D
Acq On : 09 Aug 2023 15:08
Operator : JC\MD
Sample : 03940-02DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
002-35TH-AVE(JUNE)DL

Quant Time: Aug 10 05:29:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 09 05:05:47 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/10/2023
Supervised By :Mahesh Dadoda 08/10/2023

