

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080923\
 Data File : VN078822.D
 Acq On : 09 Aug 2023 16:13
 Operator : JC\MD
 Sample : 03940-01DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(JUNE)DL

Quant Time: Aug 10 05:30:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Aug 09 05:05:47 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 08/10/2023
 Supervised By : Mahesh Dadoda 08/10/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.830	128	29364	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.112	114	151063	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.870	117	123076	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.588	65	71561	30.271	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 100.900%			
60) 4-Bromofluorobenzene	12.859	95	39872	23.080	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 76.933%			
63) Toluene-d8	10.576	98	184526	30.596	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 102.000%			
Target Compounds						
15) Acetone	4.447	58	20988	86.974	ug/l	87
25) Chloroform	7.971	83	772	0.212	ug/l #	54
30) 2-Butanone	7.494	43	5318m	4.421	ug/l	
58) 4-Methyl-2-Pentanone	10.453	43	1105	0.467	ug/l #	66
62) Toluene	10.635	91	242906	33.523	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080923\
Data File : VN078822.D
Acq On : 09 Aug 2023 16:13
Operator : JC\MD
Sample : 03940-01DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(JUNE)DL

Quant Time: Aug 10 05:30:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Aug 09 05:05:47 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/10/2023
Supervised By :Mahesh Dadoda 08/10/2023

