

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
 Data File : VN083177.D
 Acq On : 09 Aug 2024 13:10
 Operator : JC\MD
 Sample : P3393-27
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-072624

Quant Time: Aug 10 00:55:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	128604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	246521	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	249338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	112496	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	106100	57.961	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.920%
35) Dibromofluoromethane	8.165	113	84725	55.062	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.120%
50) Toluene-d8	10.565	98	313280	54.580	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.160%
62) 4-Bromofluorobenzene	12.847	95	133491	59.655	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	119.300%
Target Compounds						
20) Methylene Chloride	5.283	84	3257	1.976	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
Data File : VN083177.D
Acq On : 09 Aug 2024 13:10
Operator : JC\MD
Sample : P3393-27
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-01-072624

Quant Time: Aug 10 00:55:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

