

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
 Data File : VN083178.D
 Acq On : 09 Aug 2024 13:34
 Operator : JC\MD
 Sample : P3393-28
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-072624

Quant Time: Aug 10 00:55:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	129181	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	239568	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	243825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	107476	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	104991	57.099	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.200%
35) Dibromofluoromethane	8.171	113	81549	54.536	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.080%
50) Toluene-d8	10.565	98	311975	55.930	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.860%
62) 4-Bromofluorobenzene	12.847	95	128991	59.317	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	118.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
Data File : VN083178.D
Acq On : 09 Aug 2024 13:34
Operator : JC\MD
Sample : P3393-28
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-072624

Quant Time: Aug 10 00:55:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

