

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
 Data File : VN083185.D
 Acq On : 09 Aug 2024 16:24
 Operator : JC\MD
 Sample : P3393-21DL 5X
 Misc : 8.38g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-850-GI-SO-10-10.5-072624DL

Quant Time: Aug 10 01:00:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	124704	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	235396	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	248108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	121240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	102474	57.731	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 115.460%		
35) Dibromofluoromethane	8.171	113	79820	54.326	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 108.660%		
50) Toluene-d8	10.565	98	308761	56.335	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 112.660%		
62) 4-Bromofluorobenzene	12.847	95	137133	64.179	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 128.360%		
Target Compounds						Qvalue
44) Trichloroethene	9.353	130	9494	6.024	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
Data File : VN083185.D
Acq On : 09 Aug 2024 16:24
Operator : JC\MD
Sample : P3393-21DL 5X
Misc : 8.38g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-850-GI-SO-10-10.5-072624DL

Quant Time: Aug 10 01:00:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

