

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
 Data File : VN083186.D
 Acq On : 09 Aug 2024 16:48
 Operator : JC\MD
 Sample : P3393-24DL 5X
 Misc : 7.60g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-850-GI-SO-14.5-15-072624DL

Quant Time: Aug 10 01:01:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	125662	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	236847	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	250692	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	123005	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	104649	58.507	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 117.020%			
35) Dibromofluoromethane	8.165	113	81786	55.323	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 110.640%			
50) Toluene-d8	10.565	98	309506	56.125	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 112.240%			
62) 4-Bromofluorobenzene	12.847	95	138651	64.492	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 128.980%			
Target Compounds						
27) cis-1,2-Dichloroethene	7.477	96	2403	1.383	ug/l	76
44) Trichloroethene	9.353	130	37484	23.638	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
Data File : VN083186.D
Acq On : 09 Aug 2024 16:48
Operator : JC\MD
Sample : P3393-24DL 5X
Misc : 7.60g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-850-GI-SO-14.5-15-072624DL

Quant Time: Aug 10 01:01:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

