

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
 Data File : VN083190.D
 Acq On : 09 Aug 2024 18:26
 Operator : JC\MD
 Sample : P3393-18MEDL 100X
 Misc : 6.89g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-849-GI-SO-12.5-13-072624MEDL

Quant Time: Aug 10 01:02:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	124316	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	239520	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	113308	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	104806	59.229	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.460%
35) Dibromofluoromethane	8.165	113	80667	53.957	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.920%
50) Toluene-d8	10.570	98	313078	56.139	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	112.280%
62) 4-Bromofluorobenzene	12.847	95	131602	60.530	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	121.060%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.353	130	60081	37.466	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
Data File : VN083190.D
Acq On : 09 Aug 2024 18:26
Operator : JC\MD
Sample : P3393-18MEDL 100X
Misc : 6.89g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-849-GI-SO-12.5-13-072624MEDL

Quant Time: Aug 10 01:02:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

