

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
 Data File : VN083191.D
 Acq On : 09 Aug 2024 18:50
 Operator : JC\MD
 Sample : P3393-17MEDL 100X
 Misc : 7.71g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S-849-GI-SO-10-10.5-072624MEDL

Quant Time: Aug 10 01:03:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	125995	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	243501	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	249471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	110802	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	106997	59.662	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 119.320%	
35) Dibromofluoromethane	8.171	113	82518	54.292	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.580%	
50) Toluene-d8	10.565	98	316144	55.762	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.520%	
62) 4-Bromofluorobenzene	12.847	95	133483	60.391	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 120.780%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.383	101	2776	2.042	ug/l	90
44) Trichloroethene	9.353	130	56674	34.764	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
Data File : VN083191.D
Acq On : 09 Aug 2024 18:50
Operator : JC\MD
Sample : P3393-17MEDL 100X
Misc : 7.71g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S-849-GI-SO-10-10.5-072624MEDL

Quant Time: Aug 10 01:03:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

