

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
 Data File : VN083194.D
 Acq On : 09 Aug 2024 20:03
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Aug 10 01:04:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	126200	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	241531	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	248524	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	110576	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	106492	59.284	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 118.560%	
35) Dibromofluoromethane	8.165	113	83126	55.139	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.280%	
50) Toluene-d8	10.565	98	316707	56.317	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 112.640%	
62) 4-Bromofluorobenzene	12.847	95	133733	60.998	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 122.000%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080924\
Data File : VN083194.D
Acq On : 09 Aug 2024 20:03
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Aug 10 01:04:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

