

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081018\
 Data File : VN050520.D
 Acq On : 10 Aug 2018 16:11
 Operator : MD\SY
 Sample : PB111921ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111921ZHE#01

Quant Time: Aug 11 00:35:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	707177	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1166050	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1011212	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	403618	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	484167	51.24	ug/l	0.00
Spiked Amount			Recovery	=	102.48%	
35) Dibromofluoromethane	7.59	113	444642	47.52	ug/l	0.00
Spiked Amount			Recovery	=	95.04%	
50) Toluene-d8	10.09	98	1582291	44.98	ug/l	0.00
Spiked Amount			Recovery	=	89.96%	
62) 4-Bromofluorobenzene	12.40	95	464740	38.76	ug/l	0.00
Spiked Amount			Recovery	=	77.52%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	19919	6.72	ug/l	97
18) Methyl Acetate	4.34	43	11762	1.42	ug/l	91
20) Methylene Chloride	4.56	84	32116	2.25	ug/l	93
43) Isopropyl Acetate	8.17	43	296779	18.63	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081018\
Data File : VN050520.D
Acq On : 10 Aug 2018 16:11
Operator : MD\SY
Sample : PB111921ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111921ZHE#01

Quant Time: Aug 11 00:35:24 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

