

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081018\
 Data File : VN050525.D
 Acq On : 10 Aug 2018 18:16
 Operator : MD\SY
 Sample : PB111921ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111921ZHE#04

Quant Time: Aug 11 00:59:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	653169	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1099355	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	964076	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	385513	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	458916	52.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.18%	
35) Dibromofluoromethane	7.59	113	419702	47.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.16%	
50) Toluene-d8	10.09	98	1494570	45.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.12%	
62) 4-Bromofluorobenzene	12.40	95	444934	39.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.72%	
Target Compounds						
16) Acetone	3.83	43	20726	7.61	ug/l	96
18) Methyl Acetate	4.34	43	15235	2.07	ug/l	95
20) Methylene Chloride	4.55	84	30058	2.30	ug/l	97
43) Isopropyl Acetate	8.17	43	313626	20.89	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081018\
Data File : VN050525.D
Acq On : 10 Aug 2018 18:16
Operator : MD\SY
Sample : PB111921ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111921ZHE#04

Quant Time: Aug 11 00:59:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

