

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081018\
 Data File : VN050527.D
 Acq On : 10 Aug 2018 19:05
 Operator : MD\SY
 Sample : PB111921ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111921ZHE#06

Quant Time: Aug 11 01:10:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	654717	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1090954	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	959256	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	375995	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	449322	51.37	ug/l	0.00
Spiked Amount			Recovery	=	102.74%	
35) Dibromofluoromethane	7.60	113	413445	47.23	ug/l	0.00
Spiked Amount			Recovery	=	94.46%	
50) Toluene-d8	10.09	98	1484698	45.11	ug/l	0.00
Spiked Amount			Recovery	=	90.22%	
62) 4-Bromofluorobenzene	12.40	95	437574	39.01	ug/l	0.00
Spiked Amount			Recovery	=	78.02%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	20047	7.34	ug/l	98
18) Methyl Acetate	4.33	43	14442	1.94	ug/l	97
20) Methylene Chloride	4.55	84	29933	2.27	ug/l	95
43) Isopropyl Acetate	8.17	43	307327	20.62	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081018\
Data File : VN050527.D
Acq On : 10 Aug 2018 19:05
Operator : MD\SY
Sample : PB111921ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111921ZHE#06

Quant Time: Aug 11 01:10:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

