

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081018\
 Data File : VN050528.D
 Acq On : 10 Aug 2018 19:30
 Operator : MD\SY
 Sample : PB111921ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111921ZHE#07

Quant Time: Aug 11 01:14:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	643989	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1083201	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	933174	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	366304	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	450131	52.32	ug/l	0.00
Spiked Amount			Recovery	=	104.64%	
35) Dibromofluoromethane	7.59	113	408535	47.00	ug/l	0.00
Spiked Amount			Recovery	=	94.00%	
50) Toluene-d8	10.09	98	1475602	45.15	ug/l	0.00
Spiked Amount			Recovery	=	90.30%	
62) 4-Bromofluorobenzene	12.40	95	428089	38.44	ug/l	0.00
Spiked Amount			Recovery	=	76.88%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	19449	7.23	ug/l	97
18) Methyl Acetate	4.34	43	14952	2.06	ug/l	100
20) Methylene Chloride	4.55	84	20996	1.16	ug/l	95
43) Isopropyl Acetate	8.17	43	335068	22.65	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081018\
Data File : VN050528.D
Acq On : 10 Aug 2018 19:30
Operator : MD\SY
Sample : PB111921ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111921ZHE#07

Quant Time: Aug 11 01:14:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

