

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081018\
 Data File : VN050537.D
 Acq On : 11 Aug 2018 00:02
 Operator : MD\SY
 Sample : PB111921ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111921ZHE#11

Quant Time: Aug 11 05:44:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	631582	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1041815	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	893115	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	328941	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	437754	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
35) Dibromofluoromethane	7.59	113	404202	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
50) Toluene-d8	10.09	98	1422460	45.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.52%	
62) 4-Bromofluorobenzene	12.40	95	388759	36.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	18800	7.12	ug/l	99
18) Methyl Acetate	4.33	43	12815	1.77	ug/l	94
20) Methylene Chloride	4.56	84	19940	1.07	ug/l	87
43) Isopropyl Acetate	8.17	43	267698	18.81	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081018\
Data File : VN050537.D
Acq On : 11 Aug 2018 00:02
Operator : MD\SY
Sample : PB111921ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111921ZHE#11

Quant Time: Aug 11 05:44:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

