

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081018\
 Data File : VN050538.D
 Acq On : 11 Aug 2018 00:27
 Operator : MD\SY
 Sample : PB111921ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111921ZHE#12

Quant Time: Aug 11 05:51:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	646916	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1072405	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	941802	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	363804	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	448696	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
35) Dibromofluoromethane	7.59	113	412422	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
50) Toluene-d8	10.09	98	1487883	45.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.98%	
62) 4-Bromofluorobenzene	12.40	95	425517	38.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.18%	
Target Compounds						
16) Acetone	3.83	43	18395	6.79	ug/l	97
18) Methyl Acetate	4.33	43	14437	1.97	ug/l	97
20) Methylene Chloride	4.55	84	20052	1.03	ug/l	97
43) Isopropyl Acetate	8.17	43	278725	19.03	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081018\
Data File : VN050538.D
Acq On : 11 Aug 2018 00:27
Operator : MD\SY
Sample : PB111921ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111921ZHE#12

Quant Time: Aug 11 05:51:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

