

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081018\
 Data File : VN050542.D
 Acq On : 11 Aug 2018 2:06
 Operator : MD\SY
 Sample : PB111921ZHE#16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB111921ZHE#16

Quant Time: Aug 11 06:30:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	628444	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1065987	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	924618	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	355966	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	440141	52.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.84%	
35) Dibromofluoromethane	7.59	113	403130	47.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.26%	
50) Toluene-d8	10.09	98	1441356	44.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.64%	
62) 4-Bromofluorobenzene	12.40	95	411987	37.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.18%	
Target Compounds						
16) Acetone	3.82	43	12326	4.57	ug/l	95
18) Methyl Acetate	4.34	43	16094	2.29	ug/l #	72
43) Isopropyl Acetate	8.17	43	288165	19.79	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081018\
Data File : VN050542.D
Acq On : 11 Aug 2018 2:06
Operator : MD\SY
Sample : PB111921ZHE#16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB111921ZHE#16

Quant Time: Aug 11 06:30:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

