

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081018\
 Data File : VN050544.D
 Acq On : 11 Aug 2018 2:55
 Operator : MD\SY
 Sample : J4272-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-03-080918-B

Quant Time: Aug 11 06:34:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	623256	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1037232	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	897130	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	355665	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	432936	51.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.98%	
35) Dibromofluoromethane	7.59	113	396852	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
50) Toluene-d8	10.09	98	1420769	45.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.80%	
62) 4-Bromofluorobenzene	12.40	95	408109	38.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.54%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	19181	7.37	ug/l	95
18) Methyl Acetate	4.33	43	16603	2.39	ug/l #	70
20) Methylene Chloride	4.55	84	55320	5.92	ug/l	96
30) Chloroform	7.37	83	14954	1.10	ug/l	95
43) Isopropyl Acetate	8.17	43	265365	18.73	ug/l	96
95) Naphthalene	15.13	128	88081	8.36	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081018\
Data File : VN050544.D
Acq On : 11 Aug 2018 2:55
Operator : MD\SY
Sample : J4272-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-03-080918-B

Quant Time: Aug 11 06:34:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

