

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081018\
 Data File : VN050551.D
 Acq On : 11 Aug 2018 5:48
 Operator : MD\SY
 Sample : J4272-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-03-080918-I

Quant Time: Aug 11 06:49:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	613593	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1015830	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	843549	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	288122	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	420251	51.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.52%	
35) Dibromofluoromethane	7.59	113	385961	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
50) Toluene-d8	10.09	98	1380328	45.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.08%	
62) 4-Bromofluorobenzene	12.40	95	366584	35.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	15793	6.11	ug/l	100
18) Methyl Acetate	4.33	43	16147	2.36	ug/l	91
43) Isopropyl Acetate	8.17	43	217741	15.69	ug/l	97
95) Naphthalene	15.14	128	119030	11.53	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081018\
Data File : VN050551.D
Acq On : 11 Aug 2018 5:48
Operator : MD\SY
Sample : J4272-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
JC-03-080918-I

Quant Time: Aug 11 06:49:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

