

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081018\
 Data File : VN050554.D
 Acq On : 11 Aug 2018 7:02
 Operator : MD\SY
 Sample : J4282-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-02-080918-A

Quant Time: Aug 13 00:58:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 07 05:22:15 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	613947	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1031965	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	898413	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	344824	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	430903	52.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.06%	
35) Dibromofluoromethane	7.59	113	398100	48.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.16%	
50) Toluene-d8	10.09	98	1408214	45.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.46%	
62) 4-Bromofluorobenzene	12.40	95	404239	38.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	76.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
8) Diethyl Ether	3.41	74	4695	1.12	ug/l	83
16) Acetone	3.82	43	40242	16.11	ug/l	98
18) Methyl Acetate	4.34	43	14664	2.12	ug/l #	60
20) Methylene Chloride	4.55	84	41333	4.10	ug/l	95
30) Chloroform	7.38	83	15951	1.19	ug/l	98
43) Isopropyl Acetate	8.17	43	198027	14.05	ug/l	95
95) Naphthalene	15.13	128	10466	4.20	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081018\
Data File : VN050554.D
Acq On : 11 Aug 2018 7:02
Operator : MD\SY
Sample : J4282-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-02-080918-A

Quant Time: Aug 13 00:58:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080618W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 07 05:22:15 2018
Response via : Initial Calibration

